

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Carbazole	SOP 5-157 (GC/MS)	1.510	5.000	0.237	1.250	0.078	0.500
cis-Decalin	SOP 5-157 (GC/MS)	1.060	5.000	0.210	1.250	0.084	0.500
trans-Decalin	SOP 5-157 (GC/MS)	2.270	5.000	0.234	1.250	0.102	0.500
Benzo(b)thiophene	SOP 5-157 (GC/MS)	1.140	5.000	0.210	1.250	0.105	0.500
1,4-Dichlorobenzene	SOP 5-157 (GC/MS)	0.842	5.000	0.282	1.250	0.156	0.500
Naphthalene	SOP 5-157 (GC/MS)	4.800	5.000	0.210	1.250	0.213	0.500
2-Chloronaphthalene	SOP 5-157 (GC/MS)	0.896	5.000	0.144	1.250	0.069	0.500
2-Methylnaphthalene	SOP 5-157 (GC/MS)	2.690	5.000	0.171	1.250	0.183	0.500
1-Methylnaphthalene	SOP 5-157 (GC/MS)	3.380	5.000	0.228	1.250	0.108	0.500
2,6-dimethylnaphthalene	SOP 5-157 (GC/MS)	1.000	5.000	0.234	1.250	0.117	0.500
2,3,5-trimethylnaphthalene	SOP 5-157 (GC/MS)	0.992	5.000	0.237	1.250	0.105	0.500
Biphenyl	SOP 5-157 (GC/MS)	2.830	5.000	0.282	1.250	0.108	0.500
Acenaphthylene	SOP 5-157 (GC/MS)	1.120	5.000	0.378	1.250	0.117	0.500
Acenaphthene	SOP 5-157 (GC/MS)	1.300	5.000	0.180	1.250	0.117	0.500
Dibenzofuran	SOP 5-157 (GC/MS)	1.160	5.000	0.174	1.250	0.108	0.500
Fluorene	SOP 5-157 (GC/MS)	1.110	5.000	0.204	1.250	0.105	0.500
Anthracene	SOP 5-157 (GC/MS)	1.100	5.000	0.264	1.250	0.117	0.500
Phenanthrene	SOP 5-157 (GC/MS)	2.270	5.000	0.267	1.250	0.171	0.500
1-Methylphenanthrene	SOP 5-157 (GC/MS)	1.500	5.000	0.255	1.250	0.096	0.500
2-Methylphenanthrene	SOP 5-157 (GC/MS)	1.300	5.000	0.159	1.250	0.090	0.500
Dibenzothiophene	SOP 5-157 (GC/MS)	1.210	5.000	0.177	1.250	0.102	0.500
Fluoranthene	SOP 5-157 (GC/MS)	1.440	5.000	0.426	1.250	0.147	0.500
Pyrene	SOP 5-157 (GC/MS)	1.310	5.000	0.597	1.250	0.222	0.500
Benzo(a)anthracene	SOP 5-157 (GC/MS)	1.260	5.000	0.366	1.250	0.141	0.500
Chrysene	SOP 5-157 (GC/MS)	1.300	5.000	0.633	1.250	0.090	0.500
Benzo(b)fluoranthene	SOP 5-157 (GC/MS)	1.270	5.000	0.471	1.250	0.117	0.500
Benzo(k)fluoranthene	SOP 5-157 (GC/MS)	1.180	5.000	0.639	1.250	0.096	0.500
Benzo(e)pyrene	SOP 5-157 (GC/MS)	1.050	5.000	0.495	1.250	0.102	0.500
Benzo(a)pyrene	SOP 5-157 (GC/MS)	1.420	5.000	1.000	1.250	0.231	0.500
Perylene	SOP 5-157 (GC/MS)	1.470	5.000	0.297	1.250	0.273	0.500
Indeno(1,2,3-cd)pyrene	SOP 5-157 (GC/MS)	1.660	5.000	0.507	1.250	0.105	0.500
Dibenz(a,h)anthracene	SOP 5-157 (GC/MS)	1.320	5.000	0.474	1.250	0.093	0.500
Benzo(g,h,i)perylene	SOP 5-157 (GC/MS)	1.130	5.000	0.615	1.250	0.117	0.500
Isophorone	SOP 5-157 (GC/MS)	0.788	5.000	0.228	1.250	0.123	0.500
Retene	SOP 5-157 (GC/MS)	0.803	5.000	0.087	0.625	0.114	0.500
Benzo(b)naphtho(2,1-d)thiophene	SOP 5-157 (GC/MS)	0.738	5.000	0.066	0.625	0.081	0.500
Cholestane	SOP 5-157 (GC/MS)	1.220	5.000	0.171	0.625	0.096	0.500

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
n-Octane	5-202 (GC/FID)	48.400	500.000	2.510	50.000	9.910	50.000
n-Nonane	5-202 (GC/FID)	102.000	500.000	6.000	50.000	3.640	50.000
n-Decane	5-202 (GC/FID)	89.400	500.000	6.720	50.000	4.520	50.000
n-Undecane	5-202 (GC/FID)	83.300	500.000	7.850	50.000	4.340	50.000
n-Dodecane	5-202 (GC/FID)	77.100	500.000	6.100	50.000	4.570	50.000
n-Tridecane	5-202 (GC/FID)	80.700	500.000	5.420	50.000	4.810	50.000
Isoprenoid RRT 1380	5-202 (GC/FID)	80.700	500.000	5.420	50.000	4.810	50.000
n-Tetradecane	5-202 (GC/FID)	83.800	500.000	6.310	50.000	5.210	50.000
Isoprenoid RRT 1470	5-202 (GC/FID)	83.800	500.000	6.310	50.000	5.210	50.000
n-Pentadecane	5-202 (GC/FID)	100.000	500.000	7.160	50.000	10.400	50.000
n-Hexadecane	5-202 (GC/FID)	93.100	500.000	7.000	50.000	6.150	50.000
Norpristane (1650)	5-202 (GC/FID)	93.100	500.000	7.000	50.000	6.150	50.000
n-Heptadecane	5-202 (GC/FID)	132.000	500.000	6.940	50.000	9.060	50.000
Pristane	5-202 (GC/FID)	79.300	500.000	7.200	50.000	5.760	50.000
n-Octadecane	5-202 (GC/FID)	97.200	500.000	8.550	50.000	5.680	50.000
Phytane	5-202 (GC/FID)	85.200	500.000	7.270	50.000	4.930	50.000
n-Nonadecane	5-202 (GC/FID)	96.200	500.000	8.280	50.000	12.400	50.000
n-Eicosane	5-202 (GC/FID)	89.900	500.000	7.680	50.000	20.400	50.000
n-Heneicosane	5-202 (GC/FID)	88.800	500.000	7.530	50.000	25.800	50.000
n-Docosane	5-202 (GC/FID)	85.600	500.000	7.700	50.000	7.810	50.000
n-Tricosane	5-202 (GC/FID)	98.800	500.000	8.060	50.000	11.400	50.000
n-Tetracosane	5-202 (GC/FID)	90.200	500.000	7.140	50.000	11.100	50.000
n-Pentacosane	5-202 (GC/FID)	86.600	500.000	15.900	50.000	7.320	50.000
n-Hexacosane	5-202 (GC/FID)	85.000	500.000	7.480	50.000	8.900	50.000
n-Heptacosane	5-202 (GC/FID)	76.800	500.000	7.700	50.000	7.060	50.000
n-Octacosane	5-202 (GC/FID)	89.500	500.000	8.110	50.000	7.210	50.000
n-Nonacosane	5-202 (GC/FID)	89.900	500.000	9.310	50.000	16.200	50.000
n-Triacontane	5-202 (GC/FID)	84.700	500.000	10.200	50.000	7.270	50.000
n-Hentriacontane	5-202 (GC/FID)	77.100	500.000	8.330	50.000	16.200	50.000
n-Dotriacontane	5-202 (GC/FID)	86.700	500.000	8.220	50.000	5.390	50.000
n-Tritriacontane	5-202 (GC/FID)	83.200	500.000	8.700	50.000	4.750	50.000
n-Tetratriacontane	5-202 (GC/FID)	83.500	500.000	8.420	50.000	6.530	50.000
n-Pentatriacontane	5-202 (GC/FID)	83.000	500.000	7.920	50.000	4.800	50.000
n-Hexatriacontane	5-202 (GC/FID)	76.700	500.000	8.030	50.000	6.490	50.000
n-Heptatriacontane	5-202 (GC/FID)	79.500	500.000	7.580	50.000	4.900	50.000
n-Octatriacontane	5-202 (GC/FID)	75.800	500.000	7.910	50.000	8.100	50.000
n-Nonatriacontane	5-202 (GC/FID)	69.000	500.000	8.220	50.000	5.710	50.000

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
n-Tetracontane	5-202 (GC/FID)	54.100	500.000	7.650	50.000	15.200	50.000
TPH	5-202 (GC/FID)	32300.000	NA	2460.000	NA	5420.000	NA
n-Pentane	SOP 5-245 (GC/MS)	120.000	3560.000	1.120	14.500	NA	NA
n-Hexane	SOP 5-245 (GC/MS)	315.000	2400.000	1.920	9.780	NA	NA
n-Heptane	SOP 5-245 (GC/MS)	214.000	2480.000	2.840	10.100	NA	NA
n-Octane	SOP 5-245 (GC/MS)	206.000	2310.000	2.050	9.400	NA	NA
n-Nonane	SOP 5-245 (GC/MS)	239.000	2450.000	2.550	10.000	NA	NA
n-Decane	SOP 5-245 (GC/MS)	249.000	2270.000	2.800	9.240	NA	NA
n-Undecane	SOP 5-245 (GC/MS)	365.000	2220.000	3.730	9.060	NA	NA
Isopentane	SOP 5-245 (GC/MS)	161.000	532.000	0.870	2.170	NA	NA
2,2-Dimethylbutane	SOP 5-245 (GC/MS)	178.000	2000.000	1.150	8.150	NA	NA
2,3-Dimethylbutane	SOP 5-245 (GC/MS)	66.000	109.000	0.210	0.445	NA	NA
2-Methylpentane	SOP 5-245 (GC/MS)	83.900	867.000	1.160	3.530	NA	NA
3-Methylpentane	SOP 5-245 (GC/MS)	86.200	1420.000	0.880	5.790	NA	NA
2,2-Dimethylpentane	SOP 5-245 (GC/MS)	30.100	479.000	0.550	1.950	NA	NA
2,4-Dimethylpentane	SOP 5-245 (GC/MS)	115.000	985.000	0.830	4.010	NA	NA
2,2,3-trimethylbutane	SOP 5-245 (GC/MS)	61.400	1080.000	0.470	4.390	NA	NA
3,3-dimethylpentane	SOP 5-245 (GC/MS)	129.000	515.000	0.450	2.100	NA	NA
2-Methylhexane	SOP 5-245 (GC/MS)	61.900	634.000	0.500	2.580	NA	NA
2,3-Dimethylpentane	SOP 5-245 (GC/MS)	72.200	491.000	0.510	2.000	NA	NA
3-Methylhexane	SOP 5-245 (GC/MS)	104.000	443.000	0.590	1.800	NA	NA
3-ethylpentane	SOP 5-245 (GC/MS)	22.400	139.000	0.170	0.567	NA	NA
2,2,4-Trimethylpentane/1-heptene	SOP 5-245 (GC/MS)	292.000	3700.000	3.330	15.100	NA	NA
2,2-dimethylhexane	SOP 5-245 (GC/MS)	30.800	354.000	0.430	1.440	NA	NA
2,5-Dimethylhexane/2,2,3-trimethylpentane	SOP 5-245 (GC/MS)	108.000	1470.000	1.300	6.000	NA	NA
2,4-Dimethylhexane	SOP 5-245 (GC/MS)	81.000	455.000	0.650	1.850	NA	NA
2,3,4-Trimethylpentane	SOP 5-245 (GC/MS)	122.000	2020.000	1.370	8.220	NA	NA
2,3,3-Trimethylpentane	SOP 5-245 (GC/MS)	98.800	2010.000	1.340	8.190	NA	NA
2,3-Dimethylhexane	SOP 5-245 (GC/MS)	51.600	417.000	0.410	1.700	NA	NA
2-Methylheptane	SOP 5-245 (GC/MS)	82.800	1180.000	0.550	4.830	NA	NA
4-Methylheptane	SOP 5-245 (GC/MS)	81.500	876.000	0.710	3.570	NA	NA
3-Methylheptane	SOP 5-245 (GC/MS)	93.800	1490.000	1.550	6.090	NA	NA
3-ethylhexane	SOP 5-245 (GC/MS)	75.900	184.000	0.230	0.748	NA	NA
2,2-Dimethylheptane	SOP 5-245 (GC/MS)	109.000	1600.000	1.190	6.520	NA	NA
2,2,4-Trimethylhexane	SOP 5-245 (GC/MS)	184.000	1990.000	1.620	8.120	NA	NA
2,4-Dimethylheptane	SOP 5-245 (GC/MS)	186.000	2020.000	2.220	8.230	NA	NA
2,6-Dimethylheptane	SOP 5-245 (GC/MS)	132.000	1600.000	1.520	6.520	NA	NA

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
2,5-dimethylheptane	SOP 5-245 (GC/MS)	102.000	1490.000	1.180	6.060	NA	NA
3,3-dimethylheptane/3,5-dimethylheptane	SOP 5-245 (GC/MS)	53.200	668.000	0.700	2.720	NA	NA
2,3-Dimethylheptane	SOP 5-245 (GC/MS)	31.400	397.000	0.440	1.620	NA	NA
3,4-Dimethylheptane (D)	SOP 5-245 (GC/MS)	200.000	416.000	0.520	1.700	NA	NA
3,4-Dimethylheptane (L)	SOP 5-245 (GC/MS)	212.000	496.000	0.550	2.020	NA	NA
4-Methyloctane	SOP 5-245 (GC/MS)	205.000	1980.000	1.890	8.070	NA	NA
2-Methyloctane	SOP 5-245 (GC/MS)	105.000	982.000	1.290	4.000	NA	NA
3-Methyloctane	SOP 5-245 (GC/MS)	113.000	1510.000	1.060	6.140	NA	NA
3,3-Diethylpentane	SOP 5-245 (GC/MS)	18.800	414.000	0.600	1.680	NA	NA
2,2-Dimethyloctane	SOP 5-245 (GC/MS)	88.300	864.000	0.780	3.520	NA	NA
3,3-Dimethyloctane	SOP 5-245 (GC/MS)	78.600	852.000	0.960	3.470	NA	NA
2,3-Dimethyloctane	SOP 5-245 (GC/MS)	93.500	1000.000	1.140	4.090	NA	NA
2-methylnonane	SOP 5-245 (GC/MS)	73.700	941.000	1.150	3.830	NA	NA
3-ethyloctane	SOP 5-245 (GC/MS)	273.000	972.000	1.650	3.960	NA	NA
3-Methylnonane	SOP 5-245 (GC/MS)	73.000	1460.000	1.820	5.950	NA	NA
Benzene	SOP 5-245 (GC/MS)	76.000	2960.000	2.640	12.100	NA	NA
Toluene	SOP 5-245 (GC/MS)	55.800	1900.000	1.730	7.740	NA	NA
Ethylbenzene	SOP 5-245 (GC/MS)	69.100	2800.000	2.700	11.400	NA	NA
m-Xylene	SOP 5-245 (GC/MS)	71.700	928.000	1.230	3.780	NA	NA
p-Xylene	SOP 5-245 (GC/MS)	78.500	1970.000	2.320	8.030	NA	NA
o-Xylene	SOP 5-245 (GC/MS)	40.900	942.000	1.070	3.840	NA	NA
Isopropylbenzene	SOP 5-245 (GC/MS)	45.000	910.000	0.920	3.710	NA	NA
Propylbenzene	SOP 5-245 (GC/MS)	48.400	1790.000	1.890	7.290	NA	NA
1-Methyl-3-Ethylbenzene	SOP 5-245 (GC/MS)	37.200	918.000	0.970	3.740	NA	NA
1-Methyl-4-Ethylbenzene	SOP 5-245 (GC/MS)	26.300	816.000	0.820	3.320	NA	NA
1,3,5-Trimethylbenzene	SOP 5-245 (GC/MS)	26.100	456.000	0.450	1.860	NA	NA
1-Methyl-2-Ethylbenzene	SOP 5-245 (GC/MS)	40.400	936.000	1.140	3.820	NA	NA
tert-butylbenzene	SOP 5-245 (GC/MS)	81.700	1870.000	2.020	7.610	NA	NA
1,2,4-Trimethylbenzene	SOP 5-245 (GC/MS)	65.000	1040.000	1.200	4.220	NA	NA
isobutylbenzene	SOP 5-245 (GC/MS)	81.500	1760.000	1.800	7.180	NA	NA
sec-Butylbenzene	SOP 5-245 (GC/MS)	91.400	906.000	0.970	3.690	NA	NA
m-cymene	SOP 5-245 (GC/MS)	31.700	447.000	0.620	1.820	NA	NA
1,2,3-Trimethylbenzene	SOP 5-245 (GC/MS)	50.400	2000.000	2.420	8.160	NA	NA
p-Cymene	SOP 5-245 (GC/MS)	20.000	426.000	0.600	1.730	NA	NA
Indane	SOP 5-245 (GC/MS)	82.300	2000.000	2.440	8.160	NA	NA
o-cymene	SOP 5-245 (GC/MS)	25.400	462.000	0.440	1.880	NA	NA
1,3-Diethylbenzene	SOP 5-245 (GC/MS)	183.000	1960.000	1.820	7.990	NA	NA

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
1-methyl-3-n-propylbenzene	SOP 5-245 (GC/MS)	28.400	857.000	0.900	3.490	NA	NA
1-methyl-4-n-propylbenzene/1,4-Diethylbenzene	SOP 5-245 (GC/MS)	169.000	2900.000	2.820	11.800	NA	NA
butyl benzene	SOP 5-245 (GC/MS)	60.900	887.000	1.160	3.610	NA	NA
1,3-Dimethyl-5-Ethylbenzene	SOP 5-245 (GC/MS)	64.200	907.000	1.000	3.700	NA	NA
1,2-Diethylbenzene	SOP 5-245 (GC/MS)	38.600	433.000	0.710	1.760	NA	NA
1-Methyl-2-n-Propylbenzene	SOP 5-245 (GC/MS)	97.800	899.000	0.940	3.660	NA	NA
1,4-dimethyl-2-ethylbenzene	SOP 5-245 (GC/MS)	55.500	922.000	1.050	3.760	NA	NA
1-ethyl-2,4-dimethylbenzene	SOP 5-245 (GC/MS)	95.900	2030.000	1.760	8.260	NA	NA
1,2-Dimethyl-4-Ethylbenzene	SOP 5-245 (GC/MS)	56.600	900.000	0.890	3.670	NA	NA
1,3-Dimethyl-2-Ethylbenzene	SOP 5-245 (GC/MS)	30.300	460.000	0.490	1.870	NA	NA
1,2-dimethyl-3-ethylbenzene	SOP 5-245 (GC/MS)	50.600	852.000	0.800	3.470	NA	NA
1,2,4,5-Tetramethylbenzene	SOP 5-245 (GC/MS)	20.500	98.100	0.130	0.400	NA	NA
2-methylbutylbenzene	SOP 5-245 (GC/MS)	34.300	457.000	0.310	1.860	NA	NA
1,2,3,5-Tetramethylbenzene	SOP 5-245 (GC/MS)	210.000	2000.000	2.030	8.160	NA	NA
1,2,3,4-Tetramethylbenzene	SOP 5-245 (GC/MS)	136.000	2000.000	2.100	8.160	NA	NA
n-Pentylbenzene	SOP 5-245 (GC/MS)	146.000	1760.000	1.360	7.190	NA	NA
tert-1-butyl-3,5-dimethylbenzene	SOP 5-245 (GC/MS)	97.100	869.000	0.690	3.540	NA	NA
tert-1-butyl-4-ethylbenzene	SOP 5-245 (GC/MS)	67.400	882.000	0.820	3.590	NA	NA
1,3,5-triethylbenzene	SOP 5-245 (GC/MS)	84.900	1800.000	2.040	7.330	NA	NA
1,2,4-triethylbenzene	SOP 5-245 (GC/MS)	73.600	501.000	0.370	2.040	NA	NA
n-hexylbenzene	SOP 5-245 (GC/MS)	668.000	1750.000	1.590	7.150	NA	NA
Cyclopentane	SOP 5-245 (GC/MS)	76.500	1600.000	0.900	6.500	NA	NA
Methylcyclopentane	SOP 5-245 (GC/MS)	79.300	1040.000	1.170	4.260	NA	NA
Cyclohexane	SOP 5-245 (GC/MS)	87.200	1830.000	1.220	7.450	NA	NA
1,1-dimethylcyclopentane	SOP 5-245 (GC/MS)	109.000	1120.000	0.780	4.550	NA	NA
Methylcyclohexane	SOP 5-245 (GC/MS)	118.000	1870.000	1.300	7.620	NA	NA
ethylcyclopentane	SOP 5-245 (GC/MS)	95.300	1160.000	0.900	4.720	NA	NA
ctc-1,2,4-Trimethylcyclopentane	SOP 5-245 (GC/MS)	54.700	529.000	0.230	2.160	NA	NA
ctc-1,2,3-Trimethylcyclopentane	SOP 5-245 (GC/MS)	41.600	509.000	0.620	2.080	NA	NA
ctt-1,2,4-Trimethylcyclopentane	SOP 5-245 (GC/MS)	79.700	1220.000	1.200	4.970	NA	NA
trans-1,4-dimethylcyclohexane	SOP 5-245 (GC/MS)	135.000	1140.000	0.750	4.660	NA	NA
1-ethyl-1-methylcyclopentane	SOP 5-245 (GC/MS)	56.900	352.000	0.380	1.430	NA	NA
trans-1,2-dimethylcyclohexane	SOP 5-245 (GC/MS)	50.200	571.000	0.560	2.330	NA	NA
ccc-1,2,3-trimethylcyclopentane	SOP 5-245 (GC/MS)	60.400	259.000	0.380	1.050	NA	NA
Isopropylcyclopentane	SOP 5-245 (GC/MS)	96.100	1150.000	1.060	4.700	NA	NA
cis-1,2-dimethylcyclohexane	SOP 5-245 (GC/MS)	108.000	1200.000	1.370	4.870	NA	NA
Ethylcyclohexane/n-propylcyclopentane	SOP 5-245 (GC/MS)	151.000	3190.000	2.210	13.000	NA	NA

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
ccc-1,3,5-trimethylcyclohexane	SOP 5-245 (GC/MS)	58.300	1120.000	1.160	4.580	NA	NA
1,1,4-Trimethylcyclohexane	SOP 5-245 (GC/MS)	86.000	1160.000	1.090	4.720	NA	NA
ctt-1,2,4-Trimethylcyclohexane	SOP 5-245 (GC/MS)	74.000	1190.000	1.290	4.840	NA	NA
ctc-1,2,4-Trimethylcyclohexane	SOP 5-245 (GC/MS)	88.300	1130.000	1.080	4.590	NA	NA
1,1,2-Trimethylcyclohexane	SOP 5-245 (GC/MS)	71.800	1110.000	1.070	4.530	NA	NA
isobutylcyclopentane	SOP 5-245 (GC/MS)	128.000	1150.000	1.130	4.700	NA	NA
Isopropylcyclohexane	SOP 5-245 (GC/MS)	94.300	1880.000	1.420	7.660	NA	NA
n-Butylcyclopentane	SOP 5-245 (GC/MS)	134.000	1180.000	0.990	4.810	NA	NA
isobutylcyclohexane	SOP 5-245 (GC/MS)	147.000	1860.000	2.080	7.580	NA	NA
tert-1-methyl-2-propylcyclohexane	SOP 5-245 (GC/MS)	73.800	1090.000	1.120	4.440	NA	NA
tert-1-methyl-2-(4MP)cyclopentane	SOP 5-245 (GC/MS)	190.000	1190.000	1.700	4.860	NA	NA
3-methyl-1-butene	SOP 5-245 (GC/MS)	125.000	486.000	0.540	1.980	NA	NA
1-Pentene	SOP 5-245 (GC/MS)	74.700	1040.000	0.490	4.250	NA	NA
2-Methyl-1-Butene	SOP 5-245 (GC/MS)	37.600	368.000	0.210	1.500	NA	NA
2-methyl-1,3-butadiene	SOP 5-245 (GC/MS)	50.100	468.000	0.420	1.900	NA	NA
trans-2-Pentene	SOP 5-245 (GC/MS)	41.000	477.000	0.370	1.940	NA	NA
cis-2-Pentene	SOP 5-245 (GC/MS)	29.500	502.000	0.260	2.050	NA	NA
4-methylpentene-1	SOP 5-245 (GC/MS)	125.000	903.000	0.730	3.680	NA	NA
1-Hexene	SOP 5-245 (GC/MS)	150.000	1920.000	1.490	7.830	NA	NA
trans-2-Hexene	SOP 5-245 (GC/MS)	49.700	481.000	0.420	1.960	NA	NA
2-Methylpentene-2	SOP 5-245 (GC/MS)	76.400	879.000	0.570	3.580	NA	NA
cis-2-Hexene	SOP 5-245 (GC/MS)	51.100	1020.000	0.820	4.170	NA	NA
2-Methyl-1-Hexene/trans-1,2-dimethylcyclopentane	SOP 5-245 (GC/MS)	122.000	1990.000	1.300	8.110	NA	NA
trans-3-Heptene	SOP 5-245 (GC/MS)	59.300	1020.000	0.920	4.170	NA	NA
cis-3-Heptene	SOP 5-245 (GC/MS)	136.000	1520.000	1.750	6.180	NA	NA
trans-2-Heptene	SOP 5-245 (GC/MS)	94.900	1040.000	0.800	4.220	NA	NA
cis-2-Heptene	SOP 5-245 (GC/MS)	107.000	1570.000	1.450	6.390	NA	NA
1-Octene	SOP 5-245 (GC/MS)	210.000	2090.000	1.690	8.500	NA	NA
trans-2-Octene	SOP 5-245 (GC/MS)	38.500	555.000	0.440	2.260	NA	NA
Cis-2-octene	SOP 5-245 (GC/MS)	87.200	1040.000	0.980	4.240	NA	NA
1-Nonene	SOP 5-245 (GC/MS)	165.000	2040.000	2.230	8.330	NA	NA
trans-3-Nonene	SOP 5-245 (GC/MS)	41.500	537.000	0.880	2.190	NA	NA
cis-3-Nonene	SOP 5-245 (GC/MS)	100.000	1040.000	1.030	4.260	NA	NA
trans-2-Nonene	SOP 5-245 (GC/MS)	62.200	280.000	0.460	1.140	NA	NA
cis-2-Nonene	SOP 5-245 (GC/MS)	58.800	742.000	0.990	3.020	NA	NA
1-Decene	SOP 5-245 (GC/MS)	218.000	2070.000	1.920	8.420	NA	NA
1,2-Dichloroethane	SOP 5-245 (GC/MS)	109.000	2000.000	1.910	8.150	NA	NA

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
1,2 Dibromoethane	SOP 5-245 (GC/MS)	80.200	2000.000	1.540	8.150	NA	NA
Methyl tert butyl ether	SOP 5-245 (GC/MS)	57.900	2000.000	1.620	8.140	NA	NA
Di-isopropyl ether	SOP 5-245 (GC/MS)	88.300	2010.000	2.040	8.170	NA	NA
Ethyl tert butyl ether	SOP 5-245 (GC/MS)	185.000	2000.000	1.780	8.150	NA	NA
Tert amyl methyl ether	SOP 5-245 (GC/MS)	93.600	2000.000	2.030	8.170	NA	NA
1,2-Dichloroethylene (trans)	SOP 5-245 (GC/MS)	98.400	2000.000	1.860	8.150	NA	NA
1,2-Dichloroethylene (cis)	SOP 5-245 (GC/MS)	132.000	1990.000	1.680	8.110	NA	NA
1,1,1-Trichloroethane	SOP 5-245 (GC/MS)	210.000	2000.000	1.590	8.160	NA	NA
Trichloroethylene	SOP 5-245 (GC/MS)	122.000	2000.000	1.660	8.160	NA	NA
Tetrachloroethylene	SOP 5-245 (GC/MS)	171.000	2000.000	1.800	8.160	NA	NA
Styrene	SOP 5-245 (GC/MS)	110.000	2000.000	2.410	8.130	NA	NA
Thiophene	SOP 5-245 (GC/MS)	159.000	2000.000	1.340	8.150	NA	NA
2-Methylthiophene	SOP 5-245 (GC/MS)	48.800	1990.000	1.910	8.090	NA	NA
3-Methylthiophene	SOP 5-245 (GC/MS)	95.300	2000.000	1.710	8.140	NA	NA
2-Ethylthiophene	SOP 5-245 (GC/MS)	48.600	2000.000	1.810	8.140	NA	NA
Benzo(b)thiophene	SOP 5-245 (GC/MS)	315.000	1990.000	1.870	8.110	NA	NA
Naphthalene	SOP 5-245 (GC/MS)	122.000	999.000	0.900	4.070	NA	NA
ethanol	SOP 5-245 (GC/MS)	15600.000	20000.000	84.700	81.500	NA	NA
t-Butanol	SOP 5-245 (GC/MS)	1510.000	25600.000	27.800	104.000	NA	NA
MMT	SOP 5-245 (GC/MS)	941.000	1600.000	9.350	6.530	NA	NA
n-Dodecane	SOP 5-245 (GC/MS)	1490.000	3760.000	13.100	15.300	NA	NA
2,4'-DDD	SOP 5-128 (GC/ECD)	0.549	1.200	0.066	0.300	0.033	0.120
2,4'-DDE	SOP 5-128 (GC/ECD)	0.525	1.200	0.072	0.300	0.018	0.120
2,4'-DDT	SOP 5-128 (GC/ECD)	0.750	1.200	0.081	0.300	0.033	0.120
4,4'-DDD	SOP 5-128 (GC/ECD)	0.501	1.200	0.081	0.300	0.021	0.120
4,4'-DDE	SOP 5-128 (GC/ECD)	0.420	1.200	0.075	0.300	0.021	0.120
4,4'-DDT	SOP 5-128 (GC/ECD)	0.597	1.200	0.123	0.300	0.024	0.120
4,4'-DDMU	SOP 5-128 (GC/ECD)	0.408	1.200	0.057	0.300	0.033	0.120
aldrin	SOP 5-128 (GC/ECD)	0.501	1.200	0.057	0.300	0.027	0.120
a-chlordane	SOP 5-128 (GC/ECD)	0.417	1.200	0.060	0.300	0.021	0.120
g-chlordane	SOP 5-128 (GC/ECD)	0.378	1.200	0.054	0.300	0.036	0.120
a-BHC	SOP 5-128 (GC/ECD)	0.447	1.200	0.054	0.300	0.045	0.120
b-BHC	SOP 5-128 (GC/ECD)	0.522	1.200	0.081	0.300	0.024	0.120
d-BHC	SOP 5-128 (GC/ECD)	0.480	1.200	0.060	0.300	0.021	0.120
Lindane	SOP 5-128 (GC/ECD)	0.636	1.200	0.087	0.300	0.033	0.120
cis-nonachlor	SOP 5-128 (GC/ECD)	0.477	1.200	0.045	0.300	0.018	0.120
trans-nonachlor	SOP 5-128 (GC/ECD)	0.411	1.200	0.057	0.300	0.027	0.120

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Chlorpyrifos	SOP 5-128 (GC/ECD)	0.510	1.200	0.090	0.300	0.027	0.120
oxychlorane	SOP 5-128 (GC/ECD)	0.624	1.200	0.129	0.300	0.063	0.120
dieldrin	SOP 5-128 (GC/ECD)	0.513	1.200	0.057	0.300	0.018	0.120
endosulfan I	SOP 5-128 (GC/ECD)	0.501	1.200	0.063	0.300	0.030	0.120
endosulfan II	SOP 5-128 (GC/ECD)	0.495	1.200	0.051	0.300	0.033	0.120
endosulfan sulfate	SOP 5-128 (GC/ECD)	0.414	1.200	0.081	0.300	0.024	0.120
endrin	SOP 5-128 (GC/ECD)	0.390	1.200	0.066	0.300	0.024	0.120
endrin aldehyde	SOP 5-128 (GC/ECD)	0.492	1.200	0.084	0.300	0.036	0.120
endrin ketone	SOP 5-128 (GC/ECD)	0.558	1.200	0.057	0.300	0.018	0.120
heptachlor	SOP 5-128 (GC/ECD)	0.504	1.200	0.099	0.300	0.021	0.120
heptachlor epoxide	SOP 5-128 (GC/ECD)	0.417	1.200	0.060	0.300	0.033	0.120
Hexachlorobenzene	SOP 5-128 (GC/ECD)	0.543	1.200	0.051	0.300	0.021	0.120
methoxychlor	SOP 5-128 (GC/ECD)	0.720	1.200	0.237	0.300	0.333	0.240
Mirex	SOP 5-128 (GC/ECD)	0.417	1.200	0.054	0.300	0.018	0.120
Cl2(8)	SOP 5-128 (GC/ECD)	0.540	1.200	0.078	0.300	0.060	0.120
Cl3(18)	SOP 5-128 (GC/ECD)	0.345	1.200	0.057	0.300	0.027	0.120
Cl3(28)	SOP 5-128 (GC/ECD)	0.399	1.200	0.045	0.300	0.033	0.120
Cl4(44)	SOP 5-128 (GC/ECD)	0.504	1.200	0.054	0.300	0.024	0.120
Cl4(49)	SOP 5-128 (GC/ECD)	0.369	1.200	0.054	0.300	0.021	0.120
Cl4(52)	SOP 5-128 (GC/ECD)	0.471	1.200	0.075	0.300	0.045	0.120
Cl4(66)	SOP 5-128 (GC/ECD)	0.567	1.200	0.066	0.300	0.042	0.120
Cl4(77)	SOP 5-128 (GC/ECD)	0.432	1.200	0.114	0.300	0.054	0.120
Cl5(87)	SOP 5-128 (GC/ECD)	0.435	1.200	0.051	0.300	0.027	0.120
Cl5(101)	SOP 5-128 (GC/ECD)	0.366	1.200	0.057	0.300	0.027	0.120
Cl5(105)	SOP 5-128 (GC/ECD)	0.564	1.200	0.054	0.300	0.024	0.120
Cl5(110)	SOP 5-128 (GC/ECD)	0.818	1.200	0.096	0.300	0.057	0.120
Cl5(118)	SOP 5-128 (GC/ECD)	0.543	1.200	0.093	0.300	0.027	0.120
Cl5(126)	SOP 5-128 (GC/ECD)	0.872	1.200	0.165	0.300	0.075	0.120
Cl6(128)	SOP 5-128 (GC/ECD)	0.393	1.200	0.051	0.300	0.030	0.120
Cl6(129)	SOP 5-128 (GC/ECD)	0.755	1.200	0.117	0.300	0.030	0.120
Cl6(138)	SOP 5-128 (GC/ECD)	0.465	1.200	0.069	0.300	0.027	0.120
Cl6(153)	SOP 5-128 (GC/ECD)	0.465	1.200	0.198	0.300	0.033	0.120
Cl6(169)	SOP 5-128 (GC/ECD)	0.468	1.200	0.078	0.300	0.021	0.120
Cl7(170)	SOP 5-128 (GC/ECD)	0.426	1.200	0.057	0.300	0.027	0.120
Cl7(180)	SOP 5-128 (GC/ECD)	0.432	1.200	0.066	0.300	0.024	0.120
Cl7(183)	SOP 5-128 (GC/ECD)	0.387	1.200	0.045	0.300	0.021	0.120
Cl7(184)	SOP 5-128 (GC/ECD)	0.543	1.200	0.123	0.300	0.048	0.120

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Cl7(187)	SOP 5-128 (GC/ECD)	0.399	1.200	0.054	0.300	0.021	0.120
Cl8(195)	SOP 5-128 (GC/ECD)	0.405	1.200	0.054	0.300	0.018	0.120
Cl9(206)	SOP 5-128 (GC/ECD)	0.438	1.200	0.051	0.300	0.018	0.120
Cl10(209)	SOP 5-128 (GC/ECD)	0.420	1.200	0.057	0.300	0.033	0.120
4,4'-DDMU	SOP 5-315 (GC/MS)	0.890	1.200	0.120	0.300	0.031	0.120
2,4'-DDE	SOP 5-315 (GC/MS)	0.264	1.200	0.054	0.300	0.025	0.120
4,4'-DDE	SOP 5-315 (GC/MS)	0.291	1.200	0.054	0.300	0.028	0.120
2,4'-DDD	SOP 5-315 (GC/MS)	0.291	1.200	0.057	0.300	0.022	0.120
4,4'-DDD	SOP 5-315 (GC/MS)	0.408	1.200	0.051	0.300	0.031	0.120
2,4'-DDT	SOP 5-315 (GC/MS)	0.366	1.200	0.093	0.300	0.057	0.120
4,4'-DDT	SOP 5-315 (GC/MS)	0.477	1.200	0.093	0.300	0.041	0.120
Biphenyl	SOP 5-315 (GC/MS)	0.621	1.250	0.177	0.313	0.072	0.125
Cl1(1)	SOP 5-315 (GC/MS)	0.264	1.250	0.099	0.313	0.042	0.125
Cl1(3)	SOP 5-315 (GC/MS)	0.462	1.250	0.156	0.313	0.045	0.125
Cl2(4)	SOP 5-315 (GC/MS)	0.285	1.250	0.066	0.313	0.036	0.125
Cl2(5)	SOP 5-315 (GC/MS)	0.474	1.250	0.102	0.313	0.045	0.125
Cl2(6)	SOP 5-315 (GC/MS)	0.018	1.250	0.039	0.313	0.021	0.125
Cl2(7)	SOP 5-315 (GC/MS)	0.573	1.250	0.162	0.313	0.033	0.125
Cl2(8)	SOP 5-315 (GC/MS)	0.423	1.250	0.240	0.313	0.045	0.125
Cl2(9)	SOP 5-315 (GC/MS)	0.423	1.250	0.186	0.313	0.099	0.125
Cl2(11)	SOP 5-315 (GC/MS)	0.555	1.250	0.150	0.313	0.063	0.125
Cl2(12)	SOP 5-315 (GC/MS)	0.534	1.250	0.162	0.313	0.045	0.125
Cl2(13)	SOP 5-315 (GC/MS)	0.489	1.250	0.084	0.313	0.090	0.125
Cl2(15)	SOP 5-315 (GC/MS)	0.417	1.250	0.150	0.313	0.048	0.125
Cl3(16)	SOP 5-315 (GC/MS)	0.384	1.250	0.195	0.313	0.072	0.125
Cl3(17)	SOP 5-315 (GC/MS)	0.270	1.250	0.129	0.313	0.027	0.125
Cl3(18)	SOP 5-315 (GC/MS)	0.327	1.250	0.084	0.313	0.039	0.125
Cl3(19)	SOP 5-315 (GC/MS)	0.327	1.250	0.141	0.313	0.036	0.125
Cl3(22)	SOP 5-315 (GC/MS)	0.384	1.250	0.147	0.313	0.048	0.125
Cl3(24)	SOP 5-315 (GC/MS)	0.219	1.250	0.075	0.313	0.036	0.125
Cl3(25)	SOP 5-315 (GC/MS)	0.444	1.250	0.204	0.313	0.039	0.125
Cl3(26)	SOP 5-315 (GC/MS)	0.477	1.250	0.069	0.313	0.054	0.125
Cl3(27)	SOP 5-315 (GC/MS)	0.504	1.250	0.084	0.313	0.036	0.125
Cl3(28)	SOP 5-315 (GC/MS)	0.360	1.250	0.159	0.313	0.045	0.125
Cl3(29)	SOP 5-315 (GC/MS)	0.450	1.250	0.081	0.313	0.075	0.125
Cl3(30)	SOP 5-315 (GC/MS)	0.354	1.250	0.123	0.313	0.030	0.125
Cl3(31)	SOP 5-315 (GC/MS)	0.519	1.250	0.087	0.313	0.042	0.125

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Cl3(32)	SOP 5-315 (GC/MS)	0.375	1.250	0.123	0.313	0.036	0.125
Cl3(33)	SOP 5-315 (GC/MS)	0.318	1.250	0.156	0.313	0.030	0.125
Cl3(37)	SOP 5-315 (GC/MS)	0.609	1.250	0.195	0.313	0.069	0.125
Cl4(40)	SOP 5-315 (GC/MS)	0.447	1.250	0.222	0.313	0.084	0.125
Cl4(41)	SOP 5-315 (GC/MS)	0.564	1.250	0.189	0.313	0.051	0.125
Cl4(42)	SOP 5-315 (GC/MS)	0.423	1.250	0.147	0.313	0.039	0.125
Cl4(43)	SOP 5-315 (GC/MS)	0.402	1.250	0.171	0.313	0.048	0.125
Cl4(44)	SOP 5-315 (GC/MS)	0.285	1.250	0.111	0.313	0.048	0.125
Cl4(45)	SOP 5-315 (GC/MS)	0.324	1.250	0.096	0.313	0.039	0.125
Cl4(46)	SOP 5-315 (GC/MS)	0.318	1.250	0.135	0.313	0.033	0.125
Cl4(47)	SOP 5-315 (GC/MS)	0.432	1.250	0.099	0.313	0.075	0.125
Cl4(48)	SOP 5-315 (GC/MS)	0.381	1.250	0.129	0.313	0.066	0.125
Cl4(49)	SOP 5-315 (GC/MS)	0.189	1.250	0.162	0.313	0.069	0.125
Cl4(50)	SOP 5-315 (GC/MS)	0.282	1.250	0.075	0.313	0.030	0.125
Cl4(51)	SOP 5-315 (GC/MS)	0.279	1.250	0.072	0.313	0.027	0.125
Cl4(52)	SOP 5-315 (GC/MS)	0.507	1.250	0.150	0.313	0.042	0.125
Cl4(53)	SOP 5-315 (GC/MS)	0.285	1.250	0.084	0.313	0.033	0.125
Cl4(54)	SOP 5-315 (GC/MS)	0.240	1.250	0.114	0.313	0.024	0.125
Cl4(56)	SOP 5-315 (GC/MS)	0.405	1.250	0.096	0.313	0.042	0.125
Cl4(60)	SOP 5-315 (GC/MS)	0.609	1.250	0.114	0.313	0.051	0.125
Cl4(63)	SOP 5-315 (GC/MS)	0.633	1.250	0.135	0.313	0.054	0.125
Cl4(64)	SOP 5-315 (GC/MS)	0.935	1.250	0.114	0.313	0.060	0.125
Cl4(66)	SOP 5-315 (GC/MS)	0.675	1.250	0.147	0.313	0.042	0.125
Cl4(67)	SOP 5-315 (GC/MS)	0.663	1.250	0.078	0.313	0.045	0.125
Cl4(70)	SOP 5-315 (GC/MS)	0.420	1.250	0.156	0.313	0.087	0.125
Cl4(71)	SOP 5-315 (GC/MS)	0.273	1.250	0.081	0.313	0.054	0.125
Cl4(74)	SOP 5-315 (GC/MS)	0.570	1.250	0.126	0.313	0.072	0.125
Cl4(75)	SOP 5-315 (GC/MS)	0.192	1.250	0.129	0.313	0.042	0.125
Cl4(77)	SOP 5-315 (GC/MS)	0.426	1.250	0.150	0.313	0.060	0.125
Cl4(80)	SOP 5-315 (GC/MS)	0.621	1.250	0.093	0.313	0.066	0.125
Cl4(81)	SOP 5-315 (GC/MS)	0.462	1.250	0.093	0.313	0.096	0.125
Cl5(82)	SOP 5-315 (GC/MS)	0.633	1.250	0.093	0.313	0.045	0.125
Cl5(83)	SOP 5-315 (GC/MS)	0.564	1.250	0.099	0.313	0.036	0.125
Cl5(84)	SOP 5-315 (GC/MS)	0.486	1.250	0.153	0.313	0.054	0.125
Cl5(85)	SOP 5-315 (GC/MS)	0.741	1.250	0.258	0.313	0.072	0.125
Cl5(87)	SOP 5-315 (GC/MS)	0.417	1.250	0.087	0.313	0.093	0.125
Cl5(91)	SOP 5-315 (GC/MS)	0.294	1.250	0.135	0.313	0.036	0.125

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Cl5(92)	SOP 5-315 (GC/MS)	0.348	1.250	0.114	0.313	0.048	0.125
Cl5(95)	SOP 5-315 (GC/MS)	0.240	1.250	0.072	0.313	0.057	0.125
Cl5(97)	SOP 5-315 (GC/MS)	0.393	1.250	0.123	0.313	0.060	0.125
Cl5(99)	SOP 5-315 (GC/MS)	0.645	1.250	0.081	0.313	0.036	0.125
Cl5(100)	SOP 5-315 (GC/MS)	0.342	1.250	0.099	0.313	0.036	0.125
Cl5(101)	SOP 5-315 (GC/MS)	0.411	1.250	0.099	0.313	0.030	0.125
Cl5(104)	SOP 5-315 (GC/MS)	0.264	1.250	0.054	0.313	0.024	0.125
Cl5(105)	SOP 5-315 (GC/MS)	0.747	1.250	0.144	0.313	0.084	0.125
Cl5(110)	SOP 5-315 (GC/MS)	0.408	1.250	0.108	0.313	0.045	0.125
Cl5(114)	SOP 5-315 (GC/MS)	0.621	1.250	0.099	0.313	0.045	0.125
Cl5(115)	SOP 5-315 (GC/MS)	0.423	1.250	0.159	0.313	0.042	0.125
Cl5(118)	SOP 5-315 (GC/MS)	0.645	1.250	0.126	0.313	0.039	0.125
Cl5(123)	SOP 5-315 (GC/MS)	0.477	1.250	0.123	0.313	0.075	0.125
Cl5(124)	SOP 5-315 (GC/MS)	0.372	1.250	0.078	0.313	0.039	0.125
Cl5(125)	SOP 5-315 (GC/MS)	0.360	1.250	0.114	0.313	0.048	0.125
Cl5(126)	SOP 5-315 (GC/MS)	0.624	1.250	0.126	0.313	0.135	0.250
Cl5(127)	SOP 5-315 (GC/MS)	0.729	1.250	0.252	0.313	0.066	0.125
Cl6(128)	SOP 5-315 (GC/MS)	0.561	1.250	0.120	0.313	0.060	0.125
Cl6(130)	SOP 5-315 (GC/MS)	0.555	1.250	0.111	0.313	0.036	0.125
Cl6(131)	SOP 5-315 (GC/MS)	0.342	1.250	0.072	0.313	0.033	0.125
Cl6(134)	SOP 5-315 (GC/MS)	0.351	1.250	0.114	0.313	0.027	0.125
Cl6(135)	SOP 5-315 (GC/MS)	0.423	1.250	0.075	0.313	0.036	0.125
Cl6(136)	SOP 5-315 (GC/MS)	0.318	1.250	0.060	0.313	0.024	0.125
Cl6(137)	SOP 5-315 (GC/MS)	0.384	1.250	0.168	0.313	0.051	0.125
Cl6(138)	SOP 5-315 (GC/MS)	0.504	1.250	0.150	0.313	0.051	0.125
Cl6(139)	SOP 5-315 (GC/MS)	0.255	1.250	0.159	0.313	0.033	0.125
Cl6(140)	SOP 5-315 (GC/MS)	0.210	1.250	0.123	0.313	0.033	0.125
Cl6(141)	SOP 5-315 (GC/MS)	0.594	1.250	0.084	0.313	0.036	0.125
Cl6(144)	SOP 5-315 (GC/MS)	0.384	1.250	0.090	0.313	0.027	0.125
Cl6(146)	SOP 5-315 (GC/MS)	0.510	1.250	0.186	0.313	0.090	0.125
Cl6(149)	SOP 5-315 (GC/MS)	0.297	1.250	0.090	0.313	0.027	0.125
Cl6(151)	SOP 5-315 (GC/MS)	0.423	1.250	0.105	0.313	0.054	0.125
Cl6(153)	SOP 5-315 (GC/MS)	0.747	1.250	0.189	0.313	0.039	0.125
Cl6(154)	SOP 5-315 (GC/MS)	0.195	1.250	0.084	0.313	0.027	0.125
Cl6(155)	SOP 5-315 (GC/MS)	0.456	1.250	0.090	0.313	0.027	0.125
Cl6(156)	SOP 5-315 (GC/MS)	0.615	1.250	0.150	0.313	0.069	0.125
Cl6(157)	SOP 5-315 (GC/MS)	0.630	1.250	0.150	0.313	0.081	0.125

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Cl6(158)	SOP 5-315 (GC/MS)	0.827	1.250	0.084	0.313	0.036	0.125
Cl6(163)	SOP 5-315 (GC/MS)	0.378	1.250	0.111	0.313	0.081	0.125
Cl6(164)	SOP 5-315 (GC/MS)	0.525	1.250	0.066	0.313	0.051	0.125
Cl6(166)	SOP 5-315 (GC/MS)	0.570	1.250	0.057	0.313	0.042	0.125
Cl6(167)	SOP 5-315 (GC/MS)	0.441	1.250	0.501	0.626	0.072	0.125
Cl6(169)	SOP 5-315 (GC/MS)	0.483	1.250	0.105	0.313	0.096	0.125
Cl7(170)	SOP 5-315 (GC/MS)	0.815	1.250	0.111	0.313	0.063	0.125
Cl7(171)	SOP 5-315 (GC/MS)	0.480	1.250	0.093	0.313	0.033	0.125
Cl7(172)	SOP 5-315 (GC/MS)	0.471	1.250	0.075	0.313	0.036	0.125
Cl7(173)	SOP 5-315 (GC/MS)	0.483	1.250	0.105	0.313	0.069	0.125
Cl7(174)	SOP 5-315 (GC/MS)	0.534	1.250	0.138	0.313	0.039	0.125
Cl7(175)	SOP 5-315 (GC/MS)	0.153	1.250	0.075	0.313	0.033	0.125
Cl7(176)	SOP 5-315 (GC/MS)	0.498	1.250	0.066	0.313	0.027	0.125
Cl7(177)	SOP 5-315 (GC/MS)	0.339	1.250	0.090	0.313	0.033	0.125
Cl7(178)	SOP 5-315 (GC/MS)	0.483	1.250	0.132	0.313	0.048	0.125
Cl7(179)	SOP 5-315 (GC/MS)	0.315	1.250	0.090	0.313	0.033	0.125
Cl7(180)	SOP 5-315 (GC/MS)	0.375	1.250	0.162	0.313	0.165	0.250
Cl7(183)	SOP 5-315 (GC/MS)	0.522	1.250	0.078	0.313	0.054	0.125
Cl7(184)	SOP 5-315 (GC/MS)	0.273	1.250	0.063	0.313	0.024	0.125
Cl7(185)	SOP 5-315 (GC/MS)	0.462	1.250	0.099	0.313	0.027	0.125
Cl7(187)	SOP 5-315 (GC/MS)	0.303	1.250	0.045	0.313	0.030	0.125
Cl7(188)	SOP 5-315 (GC/MS)	0.336	1.250	0.072	0.313	0.039	0.125
Cl7(189)	SOP 5-315 (GC/MS)	0.408	1.250	0.105	0.313	0.051	0.125
Cl7(190)	SOP 5-315 (GC/MS)	0.468	1.250	0.102	0.313	0.045	0.125
Cl7(191)	SOP 5-315 (GC/MS)	0.441	1.250	0.129	0.313	0.045	0.125
Cl7(193)	SOP 5-315 (GC/MS)	0.546	1.250	0.063	0.313	0.042	0.125
Cl8(194)	SOP 5-315 (GC/MS)	0.408	1.250	0.192	0.313	0.081	0.125
Cl8(195)	SOP 5-315 (GC/MS)	0.516	1.250	0.081	0.313	0.039	0.125
Cl8(197)	SOP 5-315 (GC/MS)	0.183	1.250	0.075	0.313	0.027	0.125
Cl8(198)	SOP 5-315 (GC/MS)	0.294	1.250	0.099	0.313	0.048	0.125
Cl8(199)	SOP 5-315 (GC/MS)	0.510	1.250	0.147	0.313	0.027	0.125
Cl8(200)	SOP 5-315 (GC/MS)	0.375	1.250	0.099	0.313	0.078	0.125
Cl8(201)	SOP 5-315 (GC/MS)	0.318	1.250	0.075	0.313	0.036	0.125
Cl8(202)	SOP 5-315 (GC/MS)	0.393	1.250	0.066	0.313	0.027	0.125
Cl8(203)	SOP 5-315 (GC/MS)	0.396	1.250	0.096	0.313	0.036	0.125
Cl8(205)	SOP 5-315 (GC/MS)	0.417	1.250	0.099	0.313	0.045	0.125
Cl9(206)	SOP 5-315 (GC/MS)	0.411	1.250	0.156	0.313	0.060	0.125

Analyte	Method	Water		Sediment		Tissue	
		MDL	RL	MDL	RL	MDL	RL
Cl9(207)	SOP 5-315 (GC/MS)	0.405	1.250	0.072	0.313	0.036	0.125
Cl9(208)	SOP 5-315 (GC/MS)	0.378	1.250	0.078	0.313	0.051	0.125
Cl10(209)	SOP 5-315 (GC/MS)	0.300	1.250	0.072	0.313	0.060	0.125